

**Question:** What happens if I get a puncture?

**Answer:** The incidences of puncture are rare. If a puncture does occur, then only the section of cushion with the damage would be affected – the remaining three will still operate. The lifting cushion will not burst like a balloon and there is no risk of injury to the patient from a rapid descent. If the puncture is small, then the cushion would still work though there would be an audible leak. Punctures cannot be repaired due to the pressure and load and damaged units have to be replaced.

Contact us for more information and pricing.

